


Card for damage discription

Manufacturer: _____ Model: _____

Serialnr: _____

Pls mark the damages and use the back side for detailed descriptions.



For the protection of our employees pls send only not disinfected scope if the disinfection is not possible (leaks).
 Scope disinfected Yes No

distal end

- bending sheath bulging / perforated
- objective lens damaged
- angulation defect / deficient / tight
- play in angulation
- lightlens(s) damaged
- sprinkler head clogged / sharp
- insertion tube kinked / damaged

optical system

- CCD-Chip fault/no image
- Broken image bundle fibers
- Fluid inside the optical system
- lightbundle defect

universal unit

- universal cord damaged
- endoscope connector defect

control body

- angulation knob cracked / leak
- control housing cracked / leak

function

- suction deranged
- air-/water deranged
- auxillary defect
- forcep adoption deranged

Additional defects/ notes:
